

## PRODUCT DATA SHEET

### Alpha RA 8

The cooling technology of the cooling thermostats RA 8, RA 12 and RA 24 enables cooling performance through the entire temperature range of -25 up to 100 °C. Emptying the heat transfer liquid is done by a drain connection on the rear side of the devices. Bath covers and pump kits are included in the standard equipment.



Working temperature min.  
-25 °C



Working temperature max.  
100 °C

#### Technical Attributes

Working temperature min.	-25 °C
Working temperature max.	100 °C
Ambient temperature min.	5 °C
Ambient temperature max.	40 °C
Temperature stability	0.05 ±K
Application	external   External and inside the bath   Inside the bath
Heater power	1.5 kW
Cooling output at 20°C measured with ethanol	0.225 kW
Cooling output at 10°C measured with ethanol	0.19 kW
Cooling output at 0°C measured with ethanol	0.16 kW
Cooling output at -10°C measured with ethanol	0.13 kW
Cooling output at -20°C measured with ethanol	0.08 kW
Cooling output at -25°C measured with ethanol	0.03 kW
Power consumption	1.8 kW
Bath volume min.	5 L

#### Alpha RA 8

-25 °C
100 °C
5 °C
40 °C
0.05 ±K
external   External and inside the bath   Inside the bath
1.5 kW
0.225 kW
0.19 kW
0.16 kW
0.13 kW
0.08 kW
0.03 kW
1.8 kW
5 L

## PRODUCT DATA SHEET

### Alpha RA 8

#### Technical Attributes

Bath volume max.
Bath opening width
Bath opening depth
Bath depth
usable depth
Pump pressure max.
Pump flow max. (pressure)
Dimensions (WxDxH) in mm

#### Alpha RA 8

7.5 L
165 mm
177 mm
160 mm
140 mm
0.2 bar
15 L/min
235x500x605

LAUDA DR. R. WOBSER GMBH & CO. KG  
Pfarstraße 41/43 · 97922 Lauda-Königshofen  
Postfach 1251 · 97912 Lauda-Königshofen · DE

T +49 (0) 9343 503-0 · F +49 (0) 9343 503-222  
info@lauda.de · www.lauda.de  
WEEE-Reg-Nr.: DE 66 42 40 57

Kommunditgesellschaft: Sitz Lauda-Königshofen  
Registergericht Mannheim · HRB 560069

Persönlich haftende Gesellschafterin:  
LAUDA DR. R. WOBSER Verwaltungs-GmbH  
Sitz Lauda-Königshofen  
Registergericht Mannheim · HRB 560226

Geschäftsführer:  
Dr. Gunther Wobser (Vors.), Dr. Marc Stricker  
Beirat: Dr. Gerhard Wobser